

TEST REPORT

(Electronic version)



No: 200093461

VERIFICATION WEBSITE: www.gttc.net.cn

VERIFICATION CODE: DYHT-3298-44

ISSUE DATE: 2020-04-26

二维码载报告



APPLICANT: DONGGUAN FORES BIOTECHNOLOGY CO., LTD
ADDRESS: ROOM 201, NO. 1 BUILDING, NO. 5, SHIPAI INDUSTRIAL AVENUE, SHIPAI TOWN, DONGGUAN CITY, GUANGDONG PROVINCE

INFORMATION CONFIRMED BY APPLICANT:

DISPOSABLE FACE MASK

QUANTITY: THIRTY PIECES

MANUFACTURE'S NAME: DONGGUAN FORES BIOTECHNOLOGY CO., LTD

DATE RECEIVED/DATE TEST STARTED: 2020-04-24

CONCLUSION:

| | |
|---|---|
| pH VALUE | M |
| FORMALDEHYDE CONTENT | M |
| BREAKING STRENGTH OF MASK STRING AND JUNCTION BETWEEN MASK STRING AND MASK BODY | M |

NOTE: "M" -MEET THE STANDARD'S REQUIREMENT "F" -FAIL TO MEET THE STANDARD'S REQUIREMENT
"---" -NO COMMENT

REMARK:

THE DECISION INDICATORS ARE DERIVED FROM THE GROUP STANDARD REQUIRED BY CLIENT (T/GDBX 025-2020). OUR INSPECTION CAPACITY AUTHORIZED BY CMA COVERS THE INSPECTION ITEMS T/GDBX 025-2020.

THIS REPORT IS THE ENGLISH TRANSLATION VERSION OF THE REPORT 200072959.

ALL THE TESTED ITEMS ARE TESTED UNDER THE STANDARD CONDITION (EXCEPT FOR INDICATION).

COPIES OF THE REPORT ARE VALID ONLY RE-STAMPED.

THE EXPERIMENT WAS CARRIED OUT AT No. 1, ZHUJIANG ROAD, PANYU DISTRICT, GUANGZHOU, GUANGDONG, P. R. CHINA.

APPROVED BY:

ZiShan Guo SENIOR ENGINEER

郭子山




PAGE 1 OF 3

TEST REPORT

(Electronic version)

No:200093461



TEST REPORT

(Electronic version)

No:200093461

pH VALUE

(GB/T 7573-2009, 0.1mol/L KCl EXTRACTING SOLUTION)
(GB/T 7573-2009 MOD ISO 3071:2005)

6.4

REQUIREMENT
4.0~8.5
(T/GDBX 025-2020)

FORMALDEHYDE CONTENT

(GB/T 2912.1-2009, WATER EXTRACTION METHOD)
(GB/T 2912.1-2009 MOD ISO 14184-1:1998)

NOT DETECTED

REQUIREMENT
≤20 mg/kg
(T/GDBX 025-2020)

REMARK: DETECTION LIMIT 20mg/kg

BREAKING STRENGTH OF MASK STRING AND JUNCTION BETWEEN MASK STRING AND MASK BODY

(YY 0469-2011 5.4)

PASS

REQUIREMENT
≥10N
(T/GDBX 025-2020)



—End of Report—

PAGE 3 OF 3