

TEST REPORT

(Electronic version)



No:200093462

VERIFICATION WEBSITE: www.gttc.net.cn

VERIFICATION CODE: PCSW-3115-54

ISSUE DATE:2020-04-26

二维码下载报告



APPLICANT: DONGGUAN FORES BIOTECHNOLOGY CO., LTD
ADDRESS: ROOM 201, NO. 1 BUILDING, NO. 5, SHIPAI INDUSTRIAL AVENUE, SHIPAI TOWN, DONGGUAN CITY, GUANGDONG PROVINCE

INFORMATION CONFIRMED BY APPLICANT:

DISPOSABLE MEDICAL FACE MASK

QUANTITY: TEN PIECES

MANUFACTURE'S NAME: DONGGUAN FORES BIOTECHNOLOGY CO., LTD

DATE RECEIVED/DATE TEST STARTED: 2020-04-24

CONCLUSION:

COLIFORM GROUP	M
TOTAL AMOUNT OF BACTERIAL COLONY	M
TOTAL AMOUNT OF FUNGUS COLONY	M
PSEUDOMONAS AERUGINOSA	M
STAPHYLOCOCCUS AUREUS	M
HEMOLYTIC STREPTOCOCCUS	M


NOTE: "M" -MEET THE STANDARD'S REQUIREMENT "F" -FAIL TO MEET THE STANDARD'S REQUIREMENT
"---" -NO COMMENT

REMARK:

THIS REPORT IS THE ENGLISH TRANSLATION VERSION OF THE REPORT 200072960.
ALL THE TESTED ITEMS ARE TESTED UNDER THE STANDARD CONDITION (EXCEPT FOR INDICATION).
COPIES OF THE REPORT ARE VALID ONLY RE-STAMPED.
THE EXPERIMENT WAS CARRIED OUT AT No. 1, ZHUJIANG ROAD, PANYU DISTRICT, GUANGZHOU, GUANGDONG, P. R. CHINA.

APPROVED BY:

WanLi Hu ENGINEER




PAGE 1 OF 3

TEST REPORT

(Electronic version)

No:200093462



TEST REPORT

(Electronic version)

No:200093462

COLIFORM GROUP

(GB 15979-2002 ANNEX B)

NOT DETECTED

REQUIREMENT
MUST NOT BE DETECTED
(YY/T 0969-2013)

TOTAL AMOUNT OF BACTERIAL COLONY (CFU/g)

(GB 15979-2002 ANNEX B)

<20

REQUIREMENT
≤100
(YY/T 0969-2013)

TOTAL AMOUNT OF FUNGUS COLONY (CFU/g)

(GB 15979-2002 ANNEX B)

NOT DETECTED

REQUIREMENT
MUST NOT BE DETECTED
(YY/T 0969-2013)

PSEUDOMONAS AERUGINOSA

(GB 15979-2002 ANNEX B)

NOT DETECTED

REQUIREMENT
MUST NOT BE DETECTED
(YY/T 0969-2013)

STAPHYLOCOCCUS AUREUS

(GB 15979-2002 ANNEX B)

NOT DETECTED

REQUIREMENT
MUST NOT BE DETECTED
(YY/T 0969-2013)

HEMOLYTIC STREPTOCOCCUS

(GB 15979-2002 ANNEX B)

NOT DETECTED

REQUIREMENT
MUST NOT BE DETECTED
(YY/T 0969-2013)



—End of Report—

PAGE 3 OF 3