

TEST REPORT

(Electronic version)



No:200093463

VERIFICATION WEBSITE: www.gttc.net.cn

VERIFICATION CODE: PKTW-5857-44

ISSUE DATE:2020-04-26

检测报告



APPLICANT: DONGGUAN FORES BIOTECHNOLOGY CO.,LTD
ADDRESS: ROOM 201,NO.1 BUILDING,NO.5,SHIPAI INDUSTRIAL AVENUE,SHIPAI TOWN,DONGGUAN CITY, GUANGDONG PROVINCE

INFORMATION CONFIRMED BY APPLICANT:

DISPOSABLE MEDICAL FACE MASK

QUANTITY: TWENTY PIECES

MANUFACTURE'S NAME: DONGGUAN FORES BIOTECHNOLOGY CO.,LTD

DATE RECEIVED/DATE TEST STARTED: 2020-04-24

CONCLUSION:

RESISTANCE AGAINST PENETRATION BY SYNTHETIC BLOOD	M
PRESSURE DIFFERENCE	M
PARTICLE FILTRATION EFFICIENCY	M


NOTE: "M" -MEET THE STANDARD'S REQUIREMENT "F" -FAIL TO MEET THE STANDARD'S REQUIREMENT
"---" -NO COMMENT

REMARK:

THE AUTHORIZATION OF PRESSURE DIFFERENCE IS NOT RECEIVED FROM CNAS.
THIS REPORT IS THE ENGLISH TRANSLATION VERSION OF THE REPORT 200072961.
ALL THE TESTED ITEMS ARE TESTED UNDER THE STANDARD CONDITION (EXCEPT FOR INDICATION).
COPIES OF THE REPORT ARE VALID ONLY RE-STAMPED.
THE EXPERIMENT WAS CARRIED OUT AT No.1, ZHUJIANG ROAD, PANYU DISTRICT, GUANGZHOU, GUANGDONG, P. R. CHINA.

APPROVED BY:

WanLi Hu ENGINEER



PAGE 1 OF 3

TEST REPORT

(Electronic version)

No:200093463



TEST REPORT

(Electronic version)

No:200093463

RESISTANCE AGAINST PENETRATION BY SYNTHETIC BLOOD

(YY 0469-2011 5.5, SURFACE TENSION OF SYNTHETIC BLOOD:0.042N/m)

NO PENETRATION	REQUIREMENT NO PENETRATION APPEARED ON THE INSIDE OF MASK. (YY 0469-2011)
----------------	---

PRESSURE DIFFERENCE (Pa)

(YY 0469-2011 5.7)

26.5	REQUIREMENT ≤49 (YY 0469-2011)
------	--------------------------------------

PARTICLE FILTRATION EFFICIENCY (%)

(YY 0469-2011 5.6.2, AIR FLOW: 30L/min, AEROSOL: NaCl, AEROSOL CONCENTRATION: 15mg/m³, TEMP: 23.1°C, RH: 36.0%)

MINIMUM 91.28	REQUIREMENT ≥30 (YY 0469-2011)
---------------	--------------------------------------



————End of Report————

PAGE 3 OF 3